

Certificate



Certificate no.

CA 72112211 01

License Holder:
Elpac Electronics Inc.
17672 Armstrong Ave.
Irvine CA 92614
USA

Manufacturing Plant:
ICC Electronics (Dongguan) Ltd.
3Rd Row Block A, Shangyuan Road
2nd Industrial Estate, Qingxi
Dongguan, Guangdong
China

Test report no.: USA-RF 30791883 006

Client Reference: Carol Pickford

Tested to: CAN/CSA-C22.2 No. 60601-1-08

Certified Product: Medical Switch Mode Power Supply

License Fee - Units

Model Designation: MWCx0yz, MPCx0yz
(x = 050-100, y = 12-24)
z = A-Z, blank; not safety-relevant)

7

Rated Voltage: AC 100-240V, 50-60Hz
Rated Current: 1.5A
Protection Class: I
Output Ratings DC: 12-24V, 8.3A max., 100W max.

Appendix: 1, 1-11

7

Licensed Test mark:



c

Signature

Dipl.-Ing. M. Glagla
QA Certification Officer

Date of Issue
(day/mo/yr)
15/08/2011

Certificate



Certificate no.

CA 72112211 02

License Holder:

Elpac Electronics Inc.
17672 Armstrong Ave.
Irvine CA 92614
USA

Manufacturing Plant:

Elpac Electronics Shenzhen Inc.
HongHuaLing Industrial Zone,
XiLi Town
3F, SAST Building,
SAST Building Min Lida Ind. Garden
518055 Nanshan District, Shenzhen City
China

Test report no.: USA-RF 30791883 006

Client Reference: Carol Pickford

Tested to: CAN/CSA-C22.2 No. 60601-1-08

Certified Product: Medical Switch Mode Power Supply

License Fee - Units

Additional Manufacturing Plant: see above

(K761116)

Licensed Test mark:



Signature

A handwritten signature in blue ink that reads "Martin Glagla".

Dipl.-Ing. M. Glagla
QA Certification Officer

Date of Issue

(day/mo/yr)
15/08/2011