

Certificate



Certificate no.

CU 72072449 01

License Holder:

Elpac Electronics Inc.
Elpac Power Systems, Div. of
1562 Reynolds Avenue
Irvine CA 92614
USA

Manufacturing Plant:

Elpac Electronics Shenzhen
HongHuaLing Industrial Zone,
XiLi Town
5 F South,
SAST Building Min Lida Ind. Garden
518055 Nanshan District, Shenzhen City
China

Test report no.: USA-CC 30792801 002

Client Reference: Felix Ricardo

Tested to: UL 60601-1:2003 R6.03
CAN/CSA-C22.2 No. 601.1-M90

Certified Product: Medical Switch Mode Power Supply

License Fee - Units

Model Designation: 1) MWAx0yA
2) MWBx0yA
(x= 50-100, y= 12-24)

7

Rated Voltage: AC 100-240V, 50/60 Hz
Rated Current: 2.3 A
Protection Class: 1) I, 2) II
Output Ratings: DC 12 to 24V, 8.3Amps max.
(100 Watts max.)

7

Licensed Test mark:



Signatures

Date of Issue
(day/mo/yr)

16/11/2007

Stephan Schmitt
President

Dipl.-Ing. R. Behrends
QA Certification Officer